



# Conditions for delivery and release of surgical face masks

Versie 2020.05.25

Different types of face masks can be used as a means of prevention of and protection against the transmission of the SARS-CoV-2 virus: surgical face masks and respiratory protection masks (so-called "dust masks").

• **surgical or medical masks**: these masks are medical devices and the competent authority is the Federal Agency for Medicines and Health Products (FAMHP).







Surgical masks are designed to protect the patient from contact with infectious agents and particles from the medical staff. But they might also be intended to be worn by patients and other persons to reduce the risk of infections spreading, for example in epidemic or pandemic situations. Surgical face masks are class I medical devices. They need to comply with the requirements of the Royal Decree of 18/03/1999 on medical devices (transposition of the European Directive 93/42/EEC – `MDD') or with the requirements of the European Regulation 2017/745 on medical devices (`MDR').

The European harmonised standard these masks have to meet, is EN 14683 "Medical face masks – Requirements and test methods".

More information on medical devices can be found on our website.

respiratory protection masks (Type FFP2/FFP3 dust masks): these products are personal
protective equipment (PPE) and the competent authority is the Federal Public Service of Economy,
SMEs, Self-employed and Energy. More information on PPE can be found on the website of FPS
Economy.





### Conditions for delivery of surgical face masks before release is allowed

The compliance of surgical face masks with the essential requirements as described in the legislation, needs to be assessed by the manufacturer. The manufacturer (or his authorised representative) confirms the compliance of the products in a 'declaration of conformity' ('DoC') that should be available to clients. Compliant products that are not placed on the market in sterile conditions shall bear the CE marking as in the figure below. A CE certificate issued by a notified body ('NB') is not necessary for non-sterile surgical face masks.



Sterile surgical face masks need to bear a CE marking followed by a 4-digit code. See also figure below. These 4 digits identify the notified body that assessed the elements for securing and maintaining the sterile conditions. The notified body issues a CE certificate for this purpose.



e.g.

The CE marking needs to be present on the masks themselves or, when this is not possible, on the packaging/labelling of the masks. It should also appear on the sales packaging (e.g. boxes) and in the instructions for use (if present).

Considering the exceptional situation, we allow for deviations from these rules for CE marking and conformity assessment as described in <u>European Recommendation 2020/403</u> of the European Commission of 13 March 2020 on conformity assessment and market surveillance procedures in the context of the COVID-19 threat.

The sales and use of surgical face masks that do not bear the CE marking may be accepted exceptionally and only temporarily for the duration of the COVID-19 crisis, provided that an equivalent quality to that of standard EN 14683 can be demonstrated by a testing report according to another international standard that ensures an equivalent quality.

These alternatives might be:

- USA: ASTM F2100
- China: YY 0469:2011 and YY/T: 0969-2013

Conformity can be demonstrated by test reports or by a third party certificate. If sufficiently documented (certificates, test results according to an equivalent standard, accredited laboratory and possibility to link all documents to the batches or goods concerned), this may be accepted as an alternative.

Currently we find that Chinese masks are being tested according to European standards (for example the standards EN 14683, EN 149) by Chinese laboratories/testing bodies. Normally, compliance with European standards can only be confirmed by European testing labs with an accreditation to do so, but if the laboratory or inspection body figures on the following <u>list of bodies that are accredited by the Chinese Authorities</u>, their certificates will also be accepted (please also see <a href="https://www.cnas.org.cn/english/findanaccreditedbody/04/896740.shtml">https://www.cnas.org.cn/english/findanaccreditedbody/04/896740.shtml</a>).



FAMHP 2 | 6

Testing reports from laboratories that are **not on this list, will not be accepted**.

# Which documents should definitely be present?

#### - For non-sterile surgical face masks with CE marking:

EU declaration of conformity from manufacturer (or authorised representative). Test report of an accredited lab for all tests described in the standard EN 14683. (This report is currently requested as part of the verification of conformity to EU regulations).

#### - For sterile surgical face with CE marking:

EU declaration of conformity from manufacturer (or authorised representative)
CE certificate for the sterile aspects, issued by a notified body that is designated for medical devices under <a href="Directive 93/42/EC">Directive 93/42/EC</a> (MDD)or under <a href="Regulation 2017/745">Regulation 2017/745</a> (MDR)

Test report of an accredited lab for all tests described in the standard EN 14683.

Test report of an accredited laboratory for the sterility control.

(These reports are currently requested as part of the verification of conformity to EU regulations).

#### - For surgical face masks without CE-marking:

Testing reports from accredited lab/certificate from a third party Identification of the alternative standard that was used

### In the absence of such documents, the goods shall be deemed not to conform.

All documents should be linkable to the products in an unambiguous way. Pictures of the product, its packaging and labelling can be added as support in the offer.

## Conditions for delivery of respiratory protection masks (FFP2/FFP3) release is allowed

For these masks, all conditions can be found on the website of the FPS Economy using the following link.



# **ATTENTION: falsified certificates**



We bring to your attention that falsified and irrelevant certificates are circulating for surgical face masks and FFP masks, pretending to prove their conformity to European rules.

Below, some examples are presented that were recently seen by the Belgian Authorities. More information and more examples can also be found on the <u>website of the European Safety Federation</u>.







